


Technical Specification of Laryngoscope Set, Children & Adult

S.N.	Purchaser's Specifications		Bidder's Compliance Sheet		
			Yes/ No	Ref Dos Page No	Remarks
	Laryngoscope Set, Children & Adult				
	Manufacturer				
	Brand				
	Type / Model				
	Country of Origin				
1	Description of Function				
1.1	Laryngoscopy to assist endotracheal intubation of adults, children and infants during anaesthesia or resuscitation.				
2	Operational Requirements				
2.1	Battery powered laryngoscope unit.				
3	System Configuration				
3.1	Laryngoscope set for children & adult, complete unit.				
4	Technical Specifications				
4.1	It shall have large hollow, cylindrical, slightly ribbed handle.				
4.2	Handle to be made of either chromium-plated or stainless steel.				
4.3	It shall have provision to insert two batteries.				
4.4	It shall have provision for fitting various sizes and types of depressors.				
4.5	Light to be activated when depressor is engaged.				
4.6	Shall come with a set of four stainless steel depressors, with halogen bulb: Macintosh type: <ul style="list-style-type: none"> Curved Nr 2, length approx. 110 mm Curved Nr 3, length approx. 135 mm Curved Nr 4, length approx. 155 mm Miller type: <ul style="list-style-type: none"> Straight Nr 1, length approx. 100 mm 				
4.7	It shall be autoclaveable /sterilizeable.				
5	Accessories, spares and consumables				
5.1	Accessories: <ul style="list-style-type: none"> Spare bulbs: 1 nos. extra each Spare dry cell alkaline batteries: 2 nos. Durable protective plastic box or padded vinyl case: 1 no. 				
5.2	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).				
6	Operating Environment				
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.				
6.2	It shall be battery operated system.				
7	Standards and Safety Requirements				
7.1	Must submit ISO13485 or better for Medical Devices AND				
7.2	Must submit CE or USFDA approved product certificate.				
8	User Training				
8.1	Not applicable.				
9	Warranty				

Umesh
Er. Umesh Kumar Chaudhary
 NEC No.: 162-Biomedical A'



Medical Supply
मेडिकल सप्लाय

S.N.	Purchaser's Specifications	Bidder's Compliance Sheet		
9.1	Warranty for 1 year after acceptance.			
10	Maintenance Service During Warranty Period			
10.1	Standard warranty conditions are applicable.			
11	Installation and Commissioning			
11.1	Must supply preassembled unit, ready to use.			
12	Documentation			
12.1	User (Operating) manual in English.			
12.2	Service (Technical / Maintenance) manual in English.			

Umesh

Er. Umesh Kumar Chaudhary
NEC No.: 162-Biological A



प्रदेश सरकार
प्रदेश नं. ५
सामाजिक विकास मन्त्रालय
स्वास्थ्य विभाग
श्रीम अन्नमाल भैरव
खण्डेरी

[Signature]
मेडिकल सुपरिन्डेन्ट