


Technical Specification of Wheel Chair (foldable)

| S.N. | Purchaser's Specifications | Bidder's Offer | Remarks |
|----------|--|----------------|---------|
| | Wheel Chair (foldable) | | |
| | Manufacturer | | |
| | Brand | | |
| | Type / Model | | |
| | Country of Origin | | |
| 1 | Description of Function | | |
| 1.1 | Basic wheelchair for transportation of patients who are unable to stand/walk. | | |
| 2 | Operational Requirements | | |
| 2.1 | Basic foldable wheelchair for adult use. | | |
| 3 | System Configuration | | |
| 3.1 | Wheel Chair (foldable). | | |
| 4 | Technical Specifications | | |
| 4.1 | Heavy carriage mounted on 4 ball-bearing wheels. | | |
| 4.2 | Front wheels free rolling, 360 degrees swivel. | | |
| 4.3 | Both rear wheels with brake. | | |
| 4.4 | Foot lever, integrated in frame, facilitates tilting the wheelchair. | | |
| 4.5 | Two handles at the rear fit with plastic rims. | | |
| 4.6 | Swing-away foot and arm supports for easy stepping on/off. | | |
| 4.7 | Armrests seat and back are upholstered. | | |
| 4.8 | Materials: <ul style="list-style-type: none"> • High resistance to corrosion (tropical environment). • Frame: Chrome-plated tubular steel. • Upholstery: Plastic, flexible highly tear resistant, anti-static, flame retardant, disinfectant- and liquid proof, washable. • Tires: Heavy duty solid rubber. | | |
| 4.9 | Dimensions, Approx. $\pm 10\%$: <ul style="list-style-type: none"> • Overall: 450 x 500 x 870mm (d x w x h). • Back support: 500 x 400mm (w x h). • Frame, diameter: 23mm. • Wheels, diameter: Front 200mm, Rear 600mm. • Carrying capacity: Approximately 150kg. | | |
| 5 | Accessories, spares and consumables | | |
| 5.1 | All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above). | | |
| 6 | Operating Environment | | |
| 6.1 | The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc. | | |
| 7 | Standards and Safety Requirements | | |
| 7.1 | Must submit ISO 9001 or ISO 13485 AND | | |
| 7.2 | CE or USFDA approved product certificate. | | |
| 8 | User Training | | |


Unish



मेडिकल सुपरस्टिडेन्ड

| S.N. | Purchaser's Specifications | Bidder's Offer | Remarks |
|-----------|---|----------------|---------|
| 8.1 | Not applicable. | | |
| 9 | Warranty | | |
| 9.1 | Comprehensive warranty for 1 year. | | |
| 10 | Maintenance Service During Warranty Period | | |
| 10.1 | Standard warranty conditions are applicable. | | |
| 11 | Installation and Commissioning | | |
| 11.1 | Must supply preassembled unit, ready to use. | | |
| 12 | Documentation | | |
| 12.1 | User's manual shall be supplied in English. | | |

Umesh
Er. Umesh Kumar Chaudhary
 NEC No.: 162-Biological A



[Signature]
 मेडिकल सुपरिटेन्डेंट