

### Technical Specification of Bedside, Locker

S.N.	Purchaser's Specifications	Bidder's Compliance Sheet		
		Yes/ No	Ref Docs Page No.	Remarks
	<b>Bedside Locker</b>			
	<b>Manufacturer</b>			
	<b>Brand</b>			
	<b>Type / Model</b>			
	<b>Country of Origin</b>			
<b>1</b>	<b>Description of Function</b>			
1.1	A bedside locker simplifies the work of the caregiver and it enhances the comfort and autonomy of the patient in terms of accessibility, convenience and storage capacity.			
<b>2</b>	<b>Operational Requirements</b>			
2.1	All metal construction (machine pressed CRCA steel sheets) with heavy duty anti-corrosive and antirust treated epoxy powder coated finish (other finishes are NOT acceptable). Legs Mild steel tubular construction epoxy powder coated treated.			
<b>3</b>	<b>System Configuration</b>			
3.1	Bedside cabinet/locker, complete unit.			
<b>4</b>	<b>Technical Specifications</b>			
4.1	Feet to be capped with heavy duty plastic buffers.			
4.2	Overall approximate size 820mm H x 400mm W x 400mm L			
4.3	Fitted with superimposed stainless steel top. Top to have lip or edge or retaining rail to prevent items slipping off, Finish must be smooth.			
4.4	With stainless steel towel rail.			
4.5	Lockable drawer immediately beneath the top, minimum height of drawer 18 cm.			
4.6	Below the drawer space open on all four sides – min 20 cm height to the cupboard to allow access from all sides.			
4.7	Below the open space one cupboard with metal handle/knob with reversible hinge Cabinet door so that the door direction can be adjusted to open to the right or left depending on where it is to be used.			
<b>5</b>	<b>Accessories, spares and consumables</b>			
5.1	Not applicable.			
<b>6</b>	<b>Operating Environment</b>			
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.			
<b>7</b>	<b>Standards and Safety Requirements</b>			
7.1	Must submit ISO 9001 or ISO 13485:2003/AC:2007 <b>AND</b>			
7.2	CE or USFDA approved product certificate.			
<b>8</b>	<b>User Training</b>			
8.1	Not applicable.			

*Umesh*

**Er. Umesh Kumar Chaudhary**  
NEC No.: 162-Biome (cal A)




*Umesh*  
मेडिकल सुपरिटेन्डेंट

S.N.	Purchaser's Specifications	Bidder's Compliance Sheet		
9	<b>Warranty</b>			
9.1	Comprehensive warranty for 1 year.			
10	<b>Maintenance Service During Warranty Period</b>			
10.1	Standard warranty conditions are applicable.			
11	<b>Installation and Commissioning</b>			
11.1	Must supply preassembled unit, ready to use.			
12	<b>Documentation</b>			
12.1	Not applicable.			

*Umesh*

**Er. Umesh Kumar Chaudhary**  
NEC No.: 162-Biomedical 'A'



*[Signature]*  
**मेडिकल सुपरिटेण्डेन्ट**