Technical Specification of Cardiac ICU Bed (Manually Operating)

S.N.	Purchaser's Specifications	Bidder's Offer	Remarks
	Cardiac ICU Bed (Manually Operating)		
	Manufacturer		
	Brand		
	Type / Model		
1	Country of Origin		
1.1	Description of Function		
1.1	ICU Beds are required in the Intensive Care for comfort & safety of the patient and to facilitate comfortable transfer to		
	and fro emergency/OT/Wards etc. It is also required to carry		
	out point of care procedures including radiological		
	procedures at the bedside.		
2	Operational Requirements		
2.1	ICU Bed mainframe perforated heavy gauge sheet. The		
	system must be manually operable and adjustable for		
	heights, trendelenburg etc.		
3	System Configuration		
3.1	Cardiac ICU Bed (Manually Operating) with complete		
	accessories.		
4	Technical Specifications		
4.1	The base of the bed must have four sections.		
4.2	Must have X-Ray translucent back section made up of high		
1.2	pressure laminate.		
4.3	Must have X-Ray cassette holder underneath the back		
	section & must allow insertion of X-Ray cassette from either		
1.4	side of the bed.		
4.4	Base frame & support frame must be made up of Stainless		
	steel or pre-treated rust free oven baked epoxy powder coated washable paint.		
4.5	Must have manual adjustment for the following:		
	Height: 450-840mm		
	Back Section: 0- 50°		
	• Leg Section: 0-30°		
	• Trendelenburg (25° approx.)		
	Anti-Trendelenburg (15° approx.)		
4.6	Must have a manual quick release mechanism for back		
	section adjustment during emergency situation		
.7	Must be equipped with four articulated half-length tuck		
	away side rails		
1.8	Must be equipped with large castors (diameter 150mm) with		
	central braking and steering facility.		
.9	Four section mattress.		
.10	Mattress of the Bed must be made up of high density foam		
	with Anti-Microbial agent incorporated into all components		
	that assists in prohibiting growth of bacteria & fungi and		
1.1	easy to clean.		
.11	Mattress must be fully Radiolucent for ease in performing		
12	X-Rays.		
.12	Must have bumpers at all four corners, place for fixing accessories.	7	
13	Dimensions of bed:		
15	Difficusions of occ.		

wel.

Er. Umesh Kumar Chaudhary NEC No.: 162-Biomedical 'A' मंडिकल सुप्रस्टिन्डेन्द

S.N.	Turchaser's Specifications	Bidder's Offer	
	Length: 2200 -2290mm	Bluder's Offer	Remarks
	• Width: 850 -1020mm		
-	Mattress Size: appropriate as per bed size Accessories appropriate as per bed size		
5	recessories, spares and consumables		
5.1	Accessories:		
	Bed Ends, detachable: 01 pair		
	• Articulated half-length tuck-away side rolls, 04 N		
	• Telescopic IV Rod: 01 no.		
	• Mattress 12 cm thick: 01 no		
5.2	All standard accessories consumption and		
	and full materials to be in 1 1 1.		
	bidders must specify the quantity of		
	in their other (incliding items not specifical in		
6			
6.1	The system offered shall be designed to appear		
,	conditions include Climate. Temperature Humidity		
7	Standards and Safety Rediffrements		
7.1	Must submit ISO13485 for Medical David		
7.2	CE (93/42 EEC Directives) or HSEDA and 1		
	The state of the s		
3	User Training		
3.1	Must provide user training (including how to use and		
	Warranty		
.1	Comprehensive warranty for 2 years from acceptance.		
0	Transcerance Service Dilring Warranty David		
0.1	During the Warranty period supplier must an		
	corrective/breakdown maintenance whenever required.		
1	installation and Commissioning		
1.1	The bidder must arrange for the aguirment 1		
	and commissioned by certified or qualified		
	project of the communicate 1		
	paremaser in advance, in defail		
	Documentation		
.1	User (Operating) manual in English		
.2	Service (Technical / Maintenance) manual in English.		

Er. Umesh Kumar Chaudhary NEC No.: 162-Biomedical 'A'



ATERN FULL FOR