

Technical Specification of IV Stand

S.N.	Purchaser's Specifications		Bidder's Compliance Sheet		
			Yes /N o	Ref Docs Page No.	Remarks
	IV Stand				
	Manufacturer				
	Brand				
	Type / Model				
	Country of Origin				
1	Description of Function				
1.1	This IV/saline stand is used for hanging various intravenous items such as blood bags and glucose bottles.				
2	Operational Requirements				
2.1	Epoxy powder coated IV/Saline stand with castors.				
3	System Configuration				
3.1	Adjustable IV/saline stand with five legs, with double hooks and five swivels castors.				
4	Technical Specifications				
4.1	The IV stand shall be made of tubular anti-corrosive and antirust treated epoxy powder coated mild steel, with a 5 pronged base fitted on mobile on swivelling castors of approx. diameter Ø50mm. The castors must be non-rusting and non-marking.				
4.2	The stand should come with stainless steel double IV hook, height adjustable from approximately 1620mm to 2340mm, with a screw knob for height adjustment.				
5	Accessories, spares and consumables				
4.1	Not applicable.				
6	Operating Environment				
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.				
7	Standards and Safety Requirements				
7.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007.				
7.2	CE or USFDA approved product certificate.				
8	User Training				
8.1	Not applicable.				
9	Warranty				
9.1	Warranty for 1 year.				
10	Maintenance Service During Warranty Period				
10.1	Standard warranty conditions are applicable.				
11	Installation and Commissioning				
11.1	Must supply preassembled unit, ready to use.				
12	Documentation				
12.1	Not applicable.				

W. J.



[Signature]
मेडिकल सुपरस्टिडन्ट