

Technical Specification of Trolley, Instrument

S.N.	Purchaser's Specifications	Bidder's Offer	Remarks
	Trolley, Instrument		
	Manufacturer		
	Brand		
	Type / Model		
	Country of Origin		
1	Description of Function		
1.1	An instrument trolley for laying out surgical instruments in the operation theatre.		
2	Operational Requirements		
2.1	Stainless steel instrument trolley with swivel castors.		
3	System Configuration		
3.1	Instrument trolley with two shelves, railings, SS bowl, four swivels castors.		
4	Technical Specifications		
4.1	It shall be constructed fully with 304 grade stainless steel sheet and tube or better.		
4.2	Overall size: approximately 860 H x 460 W x 760 L mm		
4.3	It shall be have 2 tiers of grade 304 stainless steel shelves, top approx. at 880mm and lower shelf at 400mm.		
4.4	On three sides of shelves 20 mm upright lips/rail. Fourth side to have turned down edge		
4.5	Shall be mobile on 4 x 50mm diameter (approx.) robust 360 deg. swivel castors with non-marking grey tyres and with at least 2 diagonal castors shall have brakes		
5	Accessories, spares and consumables		
5.1	Accessories: <ul style="list-style-type: none"> • SS bow-1 1no. 		
6	Operating Environment		
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.		
7	Standards and Safety Requirements		
7.1	Must submit ISO 9001 or ISO 13485 AND		
7.2	CE or USFDA approved product certificate.		
8	User Training		
8.1	Not applicable.		
9	Warranty		
9.1	Comprehensive warranty for 1 year.		
10	Maintenance Service During Warranty Period		
10.1	Not applicable.		
11	Installation and Commissioning		
11.1	Must supply preassembled unit, ready to use.		
12	Documentation		
12.1	User's manual shall be supplied in English.		

Umesh
Er. Umesh Kumar Chaudhary
 NEC No.: 162-Biome cal A



[Signature]
ಮೆಡಿಕಲ್ ಸುಪರ್ಟೆಂಡೆಂಟ್