


Technical Specification Of Foot Step (Double)

S.N.	Purchaser's Specifications	Bidder's Offer	Remarks
	Foot Step (Double)		
	Manufacturer		
	Brand		
	Type / Model		
	Country of Origin		
1	Description of Function		
1.1	Footstool to assist patients ascending and descending examination/delivery table and beds in healthcare facilities.		
2	Operational Requirements		
2.1	Two step stairs for patient to mount on examination/delivery table and bed.		
3	System Configuration		
3.1	Foot Step (Double)		
4	Technical Specifications		
4.1	It shall be made of anti-corrosive and antirust treated epoxy powder coated steel with a tubular frame with heavy-duty washable finishes.		
4.2	Dimension: approximately 45 H x 45 W x 45 D cm.		
4.3	The foot step shall stand on all legs at the same time on a level surface.		
4.4	Top of the steps to have non-slip surface (e.g., embossed aluminium, stair grip or rubber)		
4.4	Feet to be fitted with heavy-duty rubber/plastic caps.		
5	Accessories, spares and consumables		
5.1	Not applicable.		
6	Operating Environment		
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.		
7	Standards and Safety Requirements		
7.1	Must submit ISO 9001 or ISO 13485 AND		
8	User Training		
8.1	Not applicable.		
9	Warranty		
9.1	Warranty for 1 year.		
10	Maintenance Service During Warranty Period		
10.1	Standard warranty conditions are applicable.		
11	Installation and Commissioning		
11.1	Must supply preassembled unit, ready to use.		
12	Documentation		
12.1	Not applicable.		

Umesh
Er. Umesh Kumar Chaudhary
 NEC No.: 162-Biome'ical 'A'



...
मेडिकल सुपरिन्टेन्डेन्ट