

### Technical Specification of Dispenser, Disinfectant

S.N.	Purchaser's Specifications	Bidder's Offer	Remarks
	<b>Disinfectant Dispenser</b>		
	<b>Manufacturer</b>		
	<b>Brand</b>		
	<b>Type / Model</b>		
	<b>Country of Origin</b>		
<b>1</b>	<b>Description of Function</b>		
1.1	Disinfectant dispenser for scrub-up		
<b>2</b>	<b>Operational Requirements</b>		
2.1	Disinfectant Dispenser		
<b>3</b>	<b>System Configuration</b>		
3.1	Disinfectant Dispenser		
<b>4</b>	<b>Technical Specifications</b>		
4.1	Wall mounted		
4.2	Must be compatible with 500ml refillable bottles		
4.3	Long lever arm control		
4.4	Stainless steel pump		
<b>5</b>	<b>Accessories, spares and consumables</b>		
5.1	<b>Accessories:</b> <ul style="list-style-type: none"> <li>• 2 refillable bottles 500ml each extra.</li> </ul>		
5.2	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).		
<b>6</b>	<b>Operating Environment</b>		
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.		
<b>7</b>	<b>Standards and Safety Requirements</b>		
7.1	This unit shall be certified to meet ISO9001 and/or ISO14971 and/or ISO 13485.		
<b>8</b>	<b>User Training</b>		
8.1	Not applicable.		
<b>9</b>	<b>Warranty</b>		
9.1	Comprehensive warranty for 1 year.		
<b>10</b>	<b>Maintenance Service During Warranty Period</b>		
10.1	Standard warranty conditions are applicable.		
<b>11</b>	<b>Installation and Commissioning</b>		
11.1	Supplier must accomplish proper commissioning of the equipment on site.		
<b>12</b>	<b>Documentation</b>		
12.1	User (Operating) manual in English		

*Umesh*  
**Er. Umesh Kumar Chaudhary**  
 NEC No.: 162-Biomedical 'A'



*Umesh*  
**मेडिकल सुपरिटेण्डेंट**